



alpha-Quartz on Membrane Filters

Frequency Count of Results for this Analyte



Results for this Laboratory



Summary of your results and their ratio to the mean

sample number	your result	'true' result	Difference(%)	standardised results
1	107.70	111.01	-2.99	0.970
2	107.90	116.16	-7.11	0.929
3	213.10	227.83	-6.46	0.935
4	215.60	230.01	-6.27	0.937

RPI	performance category
< 52	1
52 to 216	2
> 216	3
<i>RPI Reference Value: 120</i>	

Reference Value Used: No mean: 0.94

Results for Your Laboratory:

Number of outliers: 0

Performance Index (PI): 35

Running Performance Index (RPI):

Performance Category:

Results for All Participants:

Mean PI: 108.11

Median PI: 59

Mean RPI: 51.96

Median RPI: 39

Your laboratory is ranked out of 14 labs.

additional statistical data (see the statistical protocol paragraphs 12 to 18)

analysis of variance: Delta: -0.057058 SSW: 0.001025 Coefficient of variation:

RPI Plot for your Laboratory

