



alpha-Quartz on Membrane Filters

Frequency Count of Results for this Analyte



Results for this Laboratory



Summary of your results and their ratio to the mean

sample number	your result	'true' result	Difference(%)	standardised results	RPI	performance category
1	105.00	111.01	-5.42	0.946	< 52	1
2	107.40	116.16	-7.54	0.925	52 to 216	2
3	206.10	227.83	-9.54	0.905	> 216	3
4	212.30	230.01	-7.70	0.923		
Reference Value Used: No			mean:	0.92	<i>RPI Reference Value: 120</i>	

Results for Your Laboratory:

Number of outliers: 0

Performance Index (PI): 59

Running Performance Index (RPI):

Performance Category:

Results for All Participants:

Mean PI: 108.11

Median PI: 59

Mean RPI: 51.96

Median RPI: 39

Your laboratory is ranked out of 14 labs.

additional statistical data (see the statistical protocol paragraphs 12 to 18)

analysis of variance: Delta: -0.075483 SSW: 0.000852 Coefficient of variation:

RPI Plot for your Laboratory

